

TESTING LABORATORY

TEST REPORT



Chemical Testing

REPORT NO.

: TC/2020/ 478

PAGE | OF 1

CUSTOMERS NAME

: M/s Retractor Surgical Industries, Bania Bazar,

Barti Sheikh, Jalandhar,

CUSTOMERS

: Nil

ITEM NAME

: Rod Dia 10mm (used for Surgical Instruments)

IDENTIFICATION NO.

: TC/202012/232/1

DATE OF RECEIPT : 23-12-2020

DATE OF ISSUE

: 23-12-2020

Env. Condition .: Temp. 23.1° C, Rel. Humidity:- 49 %

Protocol: IS:8811:1998 & IS 9879:1998

DATE OF TEST PERFORMED: 23-12-2020

%	Si %	Mn %	P 0/0	S	Cr	Ni
0.11	0.53	0.50	0.021	0.017	11.82	0.47

NOTE: - The results listed refer only to the tested sample(s) and applicable parameters. Endorsement of product neither inferred nor implied. Total liability of our works is limited to the invoice amount. No liability will be accepted after the sample(s) is taken back. The sample(s) description is not verified in all cases and is given "as described" by customers, sample(s) not drawn by us and analysis conducted on "as received basis "unless specified otherwise. This report is not to be reproduced wholly or in parts and can not be used as an evidence in a court of law and shall not be used in media in

ULR TC616120000000000 U



This report refers only to sample supplied and may not be reproduced except in full without written permission from the General Manager, IMTT, Batala

Deputy Manager APPROVED BY

INSTITUTE FOR MACHINE TOOLS TECHNOLOG

(UNDP/UNIDO ASSISTED PB. GOVT. PROJECT) INSTITUTE WITH NABL APPROVED LABS

A-4, FOCAL POINT BATALA - 143505, PUNJAB

Ph: 01871-220009, 286847 Telefax: 01871-246472 E-mail: imtt@rediffmail.com, imttbatala@gmail.com

